



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: 9/11/2008 INITIALS: Dec OWNER: WASHINGTON STATE PATROL  
DISTRICT: 8 ADDRESS: SR101 MP77  
LOCATION Artic 77 COUNTY: Grays Harbor  
MAKE: Cardinal TYPE Trk Scale ACCURACY CLASS: IIII  
MODEL: 738 ID. NO: 9011-201D CAPACITY: 100000 lbs. SECTIONS: 2  
MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 8x10

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	4000	4000	0	CENTER	20000	20000	0
CENTER	8000	8000	0	SEC 1	20000	20000	0
CENTER	12000	12000	0	SEC 2	20000	20000	0
CENTER	16000	16000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Certificate of Accuracy", No. L2669-1&4

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.  
Signed by: David Cramer Date: Sept 11, 2008

State of Washington  
County of Shurston  
Signed and sworn to (or affirmed before  
me on Sept 11, 2008 by

Brenda A. Biscay  
(Signature of notary public)  
Brenda A. Biscay  
Printed Name of Notary  
My appointment expires 3.12.12

